KYOTO ELECTRONICS MANUFACTURING CO., LTD.

:

Application Note Determination of alcohol content in British Sprits by Alcohol meter

Industry Instrument

- Food and Beverage
- Alcohol meter
- Measurement method
- Resonant frequency oscillation

1. Scope

This is an example about determination of the alcohol content (vol%), specific gravity (t/t) and density (g/cm³) in British Sprits. The alcohol meter can be used enough for determination of these applications on the wine making process.

Technical note: The alcohol sample must be distilled as per the procedure specified on local regulations before the alcohol content can be measured by this alcohol meter.

2. Apparatus

- •Alcohol meter ALM-155
- •Distillation apparatus

3. Sample

•British sprits



•Pure water for rinsing

5. Procedure

- A: Alcohol content (vol%):
- 1) Take distilled sample solution into beaker
- 2) Introduce the sampling nozzle into the beaker
- 3) Press [Meas.] button on the instrument

B: Specific gravity (t/t) and Density (g/cm³):

- 1) Take sprits sample solution into beaker
- 2) Introduce the sampling nozzle into the beaker
- 3) Press [Meas.] button on the instrument

Measurement results

	Procedure A	Procedure B	
	Alcohol (vol%) at 20°C	S.G. (t/t)	Density (g/cm ³)
wi1	35.39	0.94963	0.94793
2	35.39	0.94963	0.94793
3	35.39	0.94963	0.94793
Mean	35.39	0.94963	0.94793
SD	0.00	0.00	0.00
RSD(%)	0.00	0.00	0.00



